

Data Sheet of gaZsens Ox



APPROVALS:

- BIS Certificate No: CM / L -7448685 as per IS :2148
- CCE Approval No: A / P / HQ / MH / 104 / 518 (P55545)

TECHNICAL SPECIFICATION :

Gas To Detect	Oxygen
Range	Standard: 0 - 25% L.E.L. Optional: Depending upon Gas Type
Sensor Type	Electrochemical
Sensor Life	Approximate 24 months depending upon sensor type
Response Time	T 95 < 15 Secs.
Oper. Temperature	0 - 55 deg C
Humidity	0-99% RH (Non - Condensing)
Signal (For Sensor)	Self Generating
Transmission	4 - 20 mA. (3 wire / 2 wire)
Transmitter Supply	24 vDC

Note:

✓ For quotation or any other information email at:

polutn.purvi@vsnl.com

ptpl@bom5.vsnl.net.in